





CALIBRATION CERTIFICATE
N° 12062V1-2




Calibration Information

GasMix model	AIOLOS II-CA	MFC	M12208189B
GasMix SN	12-062	Nominal flow	2000
Primary Reference	Bios 220M	Channel	1
Reference SN	113131	Gas name	AIR
Ref. certificate exp. date	17-oct-12	Gas comp.	(if available)

Linear Regression Calibration



Relative Standard Deviation (95% confidence level, 2*σ)



Conclusion

a	b	Dmin	Dmax	n (k=2)	p
20.090599	-5.509649	34.8	2003.5	5.3109	0.3983

Date	Expiration Date	Name	Signature
31-aout-12	30-aout-13	CARDELLA	

AlyTech - 3 rue Condorcet - 91260 Juvisy sur Orge - France - alytech@alytech.fr - www.alytech.fr - +33 (0)1 69 56 07 17

Calibration

- ❖ Calibration
- ❖ Multi Gas use
- ❖ Uncertainty calculations

